



1. Place air cushion on chair, turn valve counter-clockwise to open.




2. Connect silicon tube of hand pump to valve and inflate until cushion arches slightly.




3. Turn valve clockwise to close and remove hand pump (for two valves, repeat step 1-3).



4. With cushion in this over-inflated condition, place the user on the cushion in his/her normal sitting position with arm and foot rests properly placed.



5. Place your hand between cushion and the user's deepest penetrating bony prominence.



6. With your hand in place, slowly turn valve counter-clockwise to release some air to immerse the user in the cushion.